

Human IL-17A Precoated ELISpot Kit

Catalog Number: EXLH008



For Research Use Only. Not Intended for Diagnostic or Therapeutic Use.

info@exreprotein.com

Product Description

The Human IL-17A Precoated ELISpot Kit is used for detecting the frequency of human IL-17A-secreting cells. Human IL-17A is captured by the immobilized antibody coated on the PVDF plate and detected by the biotinylated IL-17A antibody followed by streptavidin-HRP and the 3-amino-9-ethylcarbazole (AEC) substrate. Read this manual in its entirety before use.

Limitations

- For Research Use Only. Not Intended for Diagnostic or Therapeutic Use.
- Variations in temperature, incubation time or technique while performing the assay can affect results.
- Variations in collecting, processing and storing samples may affect sample values.
- Do not substitute kit components with reagents from other sources or lots.

Materials Provided, Storage and Reconstitution

Store the kit at 2-8°C for up to 12 months. Do not use past kit expiration date.

Reagents	Quantity	Reconstitution
Biotinylated Antibody (100 µL)	1 vial	1:100 dilution with Dilution Buffer R (1X)
Streptavidin-HRP (100 µL)	1 vial	1:100 dilution with Dilution Buffer R (1X)
Cell Stimulant (50 tests)	1 bottle	Follow instructions on the vial.
Dilution Buffer R (1X) (10mL)	2 bottles	Ready-to-use
Washing Buffer (50X) (15 mL)	1 bottle	1:50 dilution with deionized water
AEC Dilution Buffer (10 mL)	1 bottle	Refer to AEC Preparation Table
AEC Solution I (20X) (500 µL)	1 vial	Refer to AEC Preparation Table
AEC Solution II (20X) (500 µL)	1 vial	Refer to AEC Preparation Table
AEC Solution III (200X) (50 µL)	1 vial	Refer to AEC Preparation Table
Pre-coated PVDF Plate	1 plate	Ready-to-use

Additional Required Supplies

- Cell culture medium (it is recommended to use serum-free medium or RPMI-1640 with 100 units/mL Penicillin, 100 µg/mL Streptomycin and 5-10% fetal calf serum)
- Adjustable pipettes and pipette tips
- Distilled or deionized water
- Lymphocyte separation medium
- CO₂ incubator
- Dissection microscope or automated ELISpot analyzer for counting spots

Precautions

- Store the kit according to the labels and instructions.
- Reagents should be at room temperature before use.
- Streptavidin-HRP and biotinylated antibody should be centrifuged briefly before use.
- Avoid cross-contamination: change pipette tips between samples, and between reagents.
- Use separate reservoirs for individual reagents.
- Ensure reagents are mixed evenly by shaking or stirring. Avoid foaming of reagents.
- Add all reagents in the same order to the plate to ensure consistent reaction times for each step.
- Remove remaining wash buffer from wells completely after each wash step. Blot the inverted plate on clean filter paper or paper towels before adding the next reagent.
- Wear gloves and avoid direct contact with AEC substrates.
- Avoid damaging the PVDF well membrane:
 - Do not make direct contact with pipette tips.
 - Do not remove the plastic plate base during incubation or substrate development. The base may be gently removed after completion of the substrate incubation to facilitate adequate washing of the well membrane and minimize background.

Preparation of Reagents

Washing Buffer (50X): Dilute 1:50 with deionized water to prepare 1X solution.

Cell Stimulant: Add sterile PBS to reconstitute lyophilized powder according to the label on the vial.

Biotinylated Antibody: Dilute 1:100 with Dilution Buffer R (1X).

Streptavidin-HRP: Dilute 1:100 with Dilution Buffer R (1X).

AEC Substrate Solution: Prepare for single-use, referring to the table below. The AEC Dilution Buffer, AEC Solution I (20X), AEC Solution II (20X), AEC Solution III (200X) are mixed well in a clean container at the ratio of 180:10:10:1. The AEC Substrate Solution should be used within 30 minutes of preparation.

Total Volume	AEC Dilution Buffer	AEC Solution I (20X)	AEC Solution II (20X)	AEC Solution III (200X)
1 mL	0.9 mL	50 µL	50 µL	5 µL
2 mL	1.8 mL	100 µL	100 µL	10 µL
3 mL	2.7 mL	150 µL	150 µL	15 µL
4 mL	3.6 mL	200 µL	200 µL	20 µL
5 mL	4.5 mL	250 µL	250 µL	25 µL
8 mL	7.2 mL	400 µL	400 µL	40 µL
10 mL	9 mL	500 µL	500 µL	50 µL

Day 1: Cell Activation (Aseptic Conditions)

1. Reagents and samples should be at room temperature before use. All samples and controls should be assayed at least in duplicate.
2. Add 200 μ L of culture medium to wet the PVDF membrane in each well and incubate for 5-10 minutes at room temperature.
3. Aspirate the culture medium from the wells.
4. Add 100 μ L of cell culture to the negative control wells
5. Add 100 μ L of the cell suspension to each well. The cell concentration should be adjusted according to the experimental design with respect to the mitogen or antigen.
6. Add 10 μ L of Cell Stimulant (antigen) to each well.
7. Cover the plate and incubate at 37°C, 5% CO₂ for 16-24 hours. The specific incubation time should be adjusted according to the cells and experimental design.

Day 2: Detection Steps (Aseptic Conditions Not Required)

1. Aspirate the wells. Add 200 μ L of ice-cold deionized water to the well and place in a refrigerator at 2-8°C for 10 minutes to remove the cells by lysis.
2. Decant or aspirate the liquid from the wells. Add 260 μ L of 1X Washing Buffer to each well with a 60 second soak. Remove the Washing Buffer by aspirating or decanting, invert the plate and blot it against clean paper towels. Wash the plate at total of 6 times. Complete removal of wash buffer at each step will improve performance.
3. Add 100 μ L of diluted Biotinylated Antibody to each well. Cover the plate and incubate at 37°C for 1 hour.
4. Repeat washing step.
5. Add 100 μ L of diluted Streptavidin-HRP to each well. Cover the plate and incubate at 37°C for 1 hour.
6. Repeat washing step.
7. Add 100 μ L of freshly prepared AEC Substrate Solution to each well. Protect from light and incubate at room temperature for 5-30 minutes. Monitor the spot development by checking the plate every 5-10 minutes.
8. Decant the liquid from the plate. Carefully remove the plastic plate base and wash the inside and outside of the experimental wells and the plastic plate base with deionized water 3-5 times. Shake off excess water, protect from light and allow to dry at room temperature before replacing the dry plate base.
9. Spots are counted manually using a dissection microscope or an automated ELISpot analyzer.